

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082423\
 Data File : VY015308.D
 Acq On : 24 Aug 2023 13:35
 Operator : SY/MD
 Sample : VY0824SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0824SBL01

Quant Time: Aug 25 03:17:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	154530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	262574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	240924	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	114585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	101774	54.865	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.740%
35) Dibromofluoromethane	7.716	113	83373	49.356	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.720%
50) Toluene-d8	10.179	98	309958	49.291	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.580%
62) 4-Bromofluorobenzene	12.477	95	105447	45.911	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.820%
Target Compounds						
20) Methylene Chloride	4.710	84	17123	8.171	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082423\
Data File : VY015308.D
Acq On : 24 Aug 2023 13:35
Operator : SY/MD
Sample : VY0824SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0824SBL01

Quant Time: Aug 25 03:17:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

