

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082423\
 Data File : VY015315.D
 Acq On : 24 Aug 2023 16:31
 Operator : SY/MD
 Sample : 04119-03
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB21153

Quant Time: Aug 25 03:21:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	19055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	29094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	24294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	9555	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	10704	46.796	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.600%
35) Dibromofluoromethane	7.715	113	9050	48.351	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.700%
50) Toluene-d8	10.178	98	32938	47.273	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.540%
62) 4-Bromofluorobenzene	12.483	95	9615	37.781	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082423\
Data File : VY015315.D
Acq On : 24 Aug 2023 16:31
Operator : SY/MD
Sample : 04119-03
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB21153

Quant Time: Aug 25 03:21:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

