

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082423\
 Data File : VY015328.D
 Acq On : 24 Aug 2023 21:35
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Aug 25 03:27:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	179014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	297306	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	272386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	147921	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	111113	51.707	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.420%
35) Dibromofluoromethane	7.716	113	91749	47.969	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.940%
50) Toluene-d8	10.179	98	352475	49.504	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.000%
62) 4-Bromofluorobenzene	12.483	95	129375	49.748	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082423\
Data File : VY015328.D
Acq On : 24 Aug 2023 21:35
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Aug 25 03:27:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

