

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082521\
 Data File : VY005837.D
 Acq On : 25 Aug 2021 14:06
 Operator : SY/MD
 Sample : M3528-01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PT-3-082421

Quant Time: Aug 26 03:37:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	32502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	49947	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	47559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	21877	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	18577	56.465	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 112.920%			
35) Dibromofluoromethane	7.728	113	15894	56.837	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 113.680%			
50) Toluene-d8	10.185	98	62555	53.338	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 106.680%			
62) 4-Bromofluorobenzene	12.483	95	18481	47.586	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 95.180%			
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	4937	10.187	ug/l	92
79) 2-Chlorotoluene	12.758	91	1971	1.935	ug/l	98
82) 4-Chlorotoluene	12.861	91	1568	1.488	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082521\
Data File : VY005837.D
Acq On : 25 Aug 2021 14:06
Operator : SY/MD
Sample : M3528-01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PT-3-082421

Quant Time: Aug 26 03:37:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

