

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082521\
 Data File : VY005838.D
 Acq On : 25 Aug 2021 14:29
 Operator : SY/MD
 Sample : M3521-05RE
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 124035RE

Quant Time: Aug 26 03:37:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	3488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	6105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	5703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1488	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	2166	61.347	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.700%
35) Dibromofluoromethane	7.728	113	1132	33.118	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	66.240%
50) Toluene-d8	10.185	98	6978	48.678	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.360%
62) 4-Bromofluorobenzene	12.489	95	1760	37.076	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082521\
Data File : VY005838.D
Acq On : 25 Aug 2021 14:29
Operator : SY/MD
Sample : M3521-05RE
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
124035RE

Quant Time: Aug 26 03:37:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

