

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082521\
 Data File : VY005845.D
 Acq On : 25 Aug 2021 17:12
 Operator : SY/MD
 Sample : M3528-01RE
 Misc : 5.08G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PT-3-082421RE

Quant Time: Aug 26 03:38:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	6545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	12183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	12903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	5237	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	5736	86.579	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 173.160%#	
35) Dibromofluoromethane	7.722	113	3484	51.078	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.160%	
50) Toluene-d8	10.185	98	15086	52.736	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.480%	
62) 4-Bromofluorobenzene	12.483	95	4810	50.775	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.560%	
Target Compounds						
20) Methylene Chloride	4.723	84	1501	18.899	ug/l	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082521\
Data File : VY005845.D
Acq On : 25 Aug 2021 17:12
Operator : SY/MD
Sample : M3528-01RE
Misc : 5.08G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PT-3-082421RE

Quant Time: Aug 26 03:38:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

