

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
 Data File : VY015342.D
 Acq On : 25 Aug 2023 15:04
 Operator : SY/MD
 Sample : 04161-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-25659

Quant Time: Aug 25 23:42:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	107320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	180352	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	164211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	80883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	70149	54.452	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	108.900%	
35) Dibromofluoromethane	7.716	113	57330	49.411	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	98.820%	
50) Toluene-d8	10.179	98	210331	48.697	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	97.400%	
62) 4-Bromofluorobenzene	12.483	95	74303	47.100	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	94.200%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.710	84	10763	6.946	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
Data File : VY015342.D
Acq On : 25 Aug 2023 15:04
Operator : SY/MD
Sample : 04161-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-25659

Quant Time: Aug 25 23:42:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

