

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
 Data File : VY015349.D
 Acq On : 25 Aug 2023 17:50
 Operator : SY/MD
 Sample : 04183-03
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-03-082523

Quant Time: Aug 25 23:45:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	54905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	87018	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	74548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	31756	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	30671	46.536	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.080%
35) Dibromofluoromethane	7.710	113	27255	48.686	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.380%
50) Toluene-d8	10.179	98	98463	47.248	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.500%
62) 4-Bromofluorobenzene	12.483	95	30854	40.535	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
Data File : VY015349.D
Acq On : 25 Aug 2023 17:50
Operator : SY/MD
Sample : 04183-03
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-03-082523

Quant Time: Aug 25 23:45:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

