

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
 Data File : VY015350.D
 Acq On : 25 Aug 2023 18:14
 Operator : SY/MD
 Sample : 04183-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-082523

Quant Time: Aug 25 23:46:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	38418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	58099	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	46051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	18923	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	19486	42.253	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.500%
35) Dibromofluoromethane	7.716	113	17670	47.275	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.540%
50) Toluene-d8	10.179	98	63450	45.602	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.200%
62) 4-Bromofluorobenzene	12.483	95	18084	35.584	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	71.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
Data File : VY015350.D
Acq On : 25 Aug 2023 18:14
Operator : SY/MD
Sample : 04183-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-082523

Quant Time: Aug 25 23:46:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

