

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
 Data File : VY015352.D
 Acq On : 25 Aug 2023 19:01
 Operator : SY/MD
 Sample : 04188-04
 Misc : 6.21g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-6

Quant Time: Aug 25 23:46:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	225871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	379005	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	349483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	170624	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	148666	54.831	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.660%
35) Dibromofluoromethane	7.710	113	120714	49.508	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.020%
50) Toluene-d8	10.179	98	440260	48.505	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.000%
62) 4-Bromofluorobenzene	12.477	95	155802	46.996	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
Data File : VY015352.D
Acq On : 25 Aug 2023 19:01
Operator : SY/MD
Sample : 04188-04
Misc : 6.21g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-6

Quant Time: Aug 25 23:46:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

