

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
 Data File : VY015332.D
 Acq On : 25 Aug 2023 10:30
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 25 13:44:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	0.253	0.191	24.5	66	0.00
3 P	Chloromethane	0.347	0.287	17.3	70	0.00
4 C	Vinyl Chloride	0.398	0.362	9.0#	71	0.00
5 T	Bromomethane	0.291	0.266	8.6	68	0.00
6 T	Chloroethane	0.289	0.296	-2.4	77	0.00
7 T	Trichlorofluoromethane	0.670	0.664	0.9	78	0.00
8 T	Diethyl Ether	0.263	0.286	-8.7	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.417	0.408	2.2	76	0.00
10 T	Methyl Iodide	0.409	0.414	-1.2	71	0.00
11 T	Tert butyl alcohol	0.116	0.116	0.0	88	0.04
12 CM	1,1-Dichloroethene	0.348	0.333	4.3#	73	0.00
13 T	Acrolein	0.056	0.032	42.9#	67	0.00
14 T	Allyl chloride	0.617	0.601	2.6	73	0.00
15 T	Acrylonitrile	0.178	0.209	-17.4	88	0.00
16 T	Acetone	0.196	0.231	-17.9	85	0.01
17 T	Carbon Disulfide	0.612	0.512	16.3	61	0.00
18 T	Methyl Acetate	0.719	0.864	-20.2	90	0.01
19 T	Methyl tert-butyl Ether	1.479	1.661	-12.3	84	0.01
20 T	Methylene Chloride	0.570	0.511	10.4	80	0.00
21 T	trans-1,2-Dichloroethene	0.392	0.389	0.8	75	0.00
22 T	Diisopropyl ether	1.539	1.631	-6.0	80	0.00
23 T	Vinyl Acetate	0.918	0.985	-7.3	79	0.00
24 P	1,1-Dichloroethane	0.833	0.872	-4.7	79	0.00
25 T	2-Butanone	0.278	0.324	-16.5	87	0.01
26 T	2,2-Dichloropropane	0.803	0.778	3.1	76	0.00
27 T	cis-1,2-Dichloroethene	0.536	0.565	-5.4	80	0.00
28 T	Bromochloromethane	0.383	0.376	1.8	76	0.00
29 T	Tetrahydrofuran	0.166	0.191	-15.1	87	0.00
30 C	Chloroform	0.934	1.011	-8.2#	83	0.00
31 T	Cyclohexane	0.597	0.518	13.2	70	0.00
32 T	1,1,1-Trichloroethane	0.795	0.831	-4.5	80	0.00
33 S	1,2-Dichloroethane-d4	0.600	0.685	-14.2	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.322	0.368	-14.3	84	0.00
36 T	1,1-Dichloropropene	0.356	0.344	3.4	74	0.00
37 T	Ethyl Acetate	0.332	0.378	-13.9	88	0.00
38 T	Carbon Tetrachloride	0.421	0.418	0.7	76	0.00
39 T	Methylcyclohexane	0.399	0.368	7.8	71	0.00
40 TM	Benzene	1.076	1.097	-2.0	78	0.00
41 T	Methacrylonitrile	0.171	0.180	-5.3	75	0.00
42 TM	1,2-Dichloroethane	0.391	0.431	-10.2	83	0.00
43 T	Isopropyl Acetate	0.604	0.675	-11.8	85	0.00
44 TM	Trichloroethene	0.310	0.317	-2.3	79	0.00
45 C	1,2-Dichloropropane	0.303	0.318	-5.0#	80	0.00
46 T	Dibromomethane	0.191	0.208	-8.9	82	0.00
47 T	Bromodichloromethane	0.457	0.505	-10.5	83	0.00
48 T	Methyl methacrylate	0.258	0.304	-17.8	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
 Data File : VY015332.D
 Acq On : 25 Aug 2023 10:30
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 25 13:44:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.005	-25.0	90	0.02
50 S	Toluene-d8	1.197	1.292	-7.9	80	0.00
51 T	4-Methyl-2-Pentanone	0.350	0.428	-22.3	91	0.00
52 CM	Toluene	0.709	0.732	-3.2#	79	0.00
53 T	t-1,3-Dichloropropene	0.480	0.521	-8.5	82	0.00
54 T	cis-1,3-Dichloropropene	0.505	0.540	-6.9	81	0.00
55 T	1,1,2-Trichloroethane	0.291	0.326	-12.0	85	0.00
56 T	Ethyl methacrylate	0.426	0.487	-14.3	84	0.00
57 T	1,3-Dichloropropane	0.475	0.530	-11.6	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.198	0.195	1.5	77	0.00
59 T	2-Hexanone	0.259	0.323	-24.7	93	0.00
60 T	Dibromochloromethane	0.347	0.399	-15.0	86	0.00
61 T	1,2-Dibromoethane	0.276	0.313	-13.4	86	0.00
62 S	4-Bromofluorobenzene	0.437	0.491	-12.4	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.348	0.376	-8.0	84	0.00
65 PM	Chlorobenzene	0.911	0.955	-4.8	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.406	-9.7	84	0.00
67 C	Ethyl Benzene	1.596	1.640	-2.8#	80	0.00
68 T	m/p-Xylenes	0.606	0.613	-1.2	78	0.00
69 T	o-Xylene	0.604	0.635	-5.1	82	0.00
70 T	Styrene	1.045	1.116	-6.8	82	0.00
71 P	Bromoform	0.268	0.308	-14.9	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	3.215	3.257	-1.3	82	0.00
74 T	N-amyl acetate	1.161	1.258	-8.4	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.837	0.949	-13.4	89	0.00
76 T	1,2,3-Trichloropropane	0.607	0.784	-29.2#	105	0.00
77 T	Bromobenzene	0.784	0.809	-3.2	83	0.00
78 T	n-propylbenzene	3.818	3.796	0.6	80	0.00
79 T	2-Chlorotoluene	2.203	2.219	-0.7	81	0.00
80 T	1,3,5-Trimethylbenzene	2.680	2.704	-0.9	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.277	0.301	-8.7	83	0.00
82 T	4-Chlorotoluene	2.275	2.288	-0.6	80	0.00
83 T	tert-Butylbenzene	2.451	2.507	-2.3	82	0.00
84 T	1,2,4-Trimethylbenzene	2.678	2.734	-2.1	81	0.00
85 T	sec-Butylbenzene	3.592	3.619	-0.8	81	0.00
86 T	p-Isopropyltoluene	2.978	3.008	-1.0	80	0.00
87 T	1,3-Dichlorobenzene	1.554	1.603	-3.2	82	0.00
88 T	1,4-Dichlorobenzene	1.559	1.618	-3.8	83	0.00
89 T	n-Butylbenzene	2.808	2.798	0.4	79	0.00
90 T	Hexachloroethane	0.591	0.578	2.2	79	0.00
91 T	1,2-Dichlorobenzene	1.452	1.546	-6.5	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.162	0.196	-21.0	95	0.00
93 T	1,2,4-Trichlorobenzene	0.932	1.005	-7.8	86	0.00
94 T	Hexachlorobutadiene	0.507	0.505	0.4	81	0.00
95 T	Naphthalene	2.322	2.862	-23.3	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
Data File : VY015332.D
Acq On : 25 Aug 2023 10:30
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 25 13:44:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.842	0.948	-12.6	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6