



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082523\
Data File : VY015331.D
Acq On : 25 Aug 2023 08:47
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Title : SW846 8260
Last Update : Mon Aug 14 12:43:26 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.8	21968	PASS
75	95	30	60	54.0	59880	PASS
95	95	100	100	100.0	110909	PASS
96	95	5	9	6.7	7439	PASS
173	174	0.00	2	1.1	1050	PASS
174	95	50	100	86.2	95603	PASS
175	174	5	9	7.6	7232	PASS
176	174	95	101	95.9	91720	PASS
177	176	5	9	6.5	5994	PASS