

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082621\
 Data File : VY005858.D
 Acq On : 26 Aug 2021 14:09
 Operator : SY/MD
 Sample : M3547-01
 Misc : 5.15G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-1

Quant Time: Aug 27 03:25:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	187954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	321625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	342812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	176719	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	116868	61.427	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.860%	
35) Dibromofluoromethane	7.728	113	99582	55.302	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.600%	
50) Toluene-d8	10.185	98	421423	55.803	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.600%	
62) 4-Bromofluorobenzene	12.483	95	150687	60.254	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.500%	
Target Compounds						
20) Methylene Chloride	4.729	84	24495	7.760	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082621\
Data File : VY005858.D
Acq On : 26 Aug 2021 14:09
Operator : SY/MD
Sample : M3547-01
Misc : 5.15G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-1

Quant Time: Aug 27 03:25:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

