

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082621\
 Data File : VY005861.D
 Acq On : 26 Aug 2021 15:18
 Operator : SY/MD
 Sample : M3547-04
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-3

Quant Time: Aug 27 03:26:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	191406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	330044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	347004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	179848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	119956	61.913	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.820%	
35) Dibromofluoromethane	7.728	113	100712	54.503	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.000%	
50) Toluene-d8	10.185	98	428750	55.325	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.640%	
62) 4-Bromofluorobenzene	12.483	95	151877	59.181	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 118.360%	
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	30805	11.204	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082621\
Data File : VY005861.D
Acq On : 26 Aug 2021 15:18
Operator : SY/MD
Sample : M3547-04
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-3

Quant Time: Aug 27 03:26:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

