

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082621\
 Data File : VY005865.D
 Acq On : 26 Aug 2021 16:51
 Operator : SY/MD
 Sample : M3547-08
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-7

Quant Time: Aug 27 03:26:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	22309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	33837	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	32910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	14492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	12671	56.110	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.220%
35) Dibromofluoromethane	7.722	113	10755	56.771	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.540%
50) Toluene-d8	10.185	98	43221	54.399	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.800%
62) 4-Bromofluorobenzene	12.483	95	12890	48.992	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.980%
Target Compounds						
20) Methylene Chloride	4.728	84	3202	9.245	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082621\
Data File : VY005865.D
Acq On : 26 Aug 2021 16:51
Operator : SY/MD
Sample : M3547-08
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-7

Quant Time: Aug 27 03:26:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

