



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082621\
Data File : VY005848.D
Acq On : 26 Aug 2021 09:27
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Title : SW846 8260
Last Update : Tue Aug 03 04:52:05 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	27664	PASS
75	95	30	60	53.0	74579	PASS
95	95	100	100	100.0	140720	PASS
96	95	5	9	6.5	9204	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	96.1	135264	PASS
175	174	5	9	7.7	10460	PASS
176	174	95	101	95.3	128845	PASS
177	176	5	9	6.5	8337	PASS