

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082721\
 Data File : VY005878.D
 Acq On : 27 Aug 2021 13:09
 Operator : SY/MD
 Sample : M3548-01RE
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-235RE

Quant Time: Aug 28 03:50:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	7804	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	12046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	12013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4683	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	4902	62.054	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.100%
35) Dibromofluoromethane	7.722	113	3085	45.743	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.480%
50) Toluene-d8	10.185	98	15137	53.516	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.040%
62) 4-Bromofluorobenzene	12.483	95	4587	48.972	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082721\
Data File : VY005878.D
Acq On : 27 Aug 2021 13:09
Operator : SY/MD
Sample : M3548-01RE
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-235RE

Quant Time: Aug 28 03:50:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

