

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082721\
 Data File : VY005885.D
 Acq On : 27 Aug 2021 15:51
 Operator : SY/MD
 Sample : M3569-03
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-1A

Quant Time: Aug 28 03:51:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	9706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	15651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	16559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	8339	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	5868	59.726	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.460%
35) Dibromofluoromethane	7.728	113	3737	42.647	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.300%
50) Toluene-d8	10.185	98	21045	57.266	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.540%
62) 4-Bromofluorobenzene	12.489	95	6078	49.944	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082721\
Data File : VY005885.D
Acq On : 27 Aug 2021 15:51
Operator : SY/MD
Sample : M3569-03
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-1A

Quant Time: Aug 28 03:51:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

