

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082721\
 Data File : VY005872.D
 Acq On : 27 Aug 2021 10:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 28 03:49:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	67	0.00
2 T	Dichlorodifluoromethane	0.411	0.344	16.3	62	0.00
3 P	Chloromethane	0.491	0.423	13.8	60	0.00
4 C	Vinyl Chloride	0.588	0.578	1.7#	67	0.00
5 T	Bromomethane	0.387	0.412	-6.5	74	0.00
6 T	Chloroethane	0.355	0.380	-7.0	69	0.00
7 T	Trichlorofluoromethane	0.787	0.801	-1.8	66	0.00
8 T	Diethyl Ether	0.275	0.278	-1.1	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.427	0.443	-3.7	72	0.00
10 T	Methyl Iodide	0.532	0.487	8.5	63	0.00
11 T	Tert butyl alcohol	0.050	0.066	-32.0#	111	0.00
12 CM	1,1-Dichloroethene	0.431	0.426	1.2#	69	0.00
13 T	Acrolein	0.030	0.024	20.0	55	0.00
14 T	Allyl chloride	0.755	0.776	-2.8	71	0.00
15 T	Acrylonitrile	0.141	0.151	-7.1	82	0.00
16 T	Acetone	0.128	0.150	-17.2	85	0.00
17 T	Carbon Disulfide	1.328	1.132	14.8	60	0.00
18 T	Methyl Acetate	0.334	0.343	-2.7	84	0.00
19 T	Methyl tert-butyl Ether	1.335	1.431	-7.2	77	0.00
20 T	Methylene Chloride	0.641	0.534	16.7	71	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.490	-1.7	72	0.00
22 T	Diisopropyl ether	1.557	1.685	-8.2	75	0.00
23 T	Vinyl Acetate	1.036	1.111	-7.2	77	0.00
24 P	1,1-Dichloroethane	0.863	0.942	-9.2	75	0.00
25 T	2-Butanone	0.189	0.207	-9.5	84	0.00
26 T	2,2-Dichloropropane	0.788	0.872	-10.7	76	0.00
27 T	cis-1,2-Dichloroethene	0.548	0.587	-7.1	74	0.00
28 T	Bromochloromethane	0.344	0.332	3.5	73	0.00
29 T	Tetrahydrofuran	0.125	0.130	-4.0	83	0.00
30 C	Chloroform	0.881	0.978	-11.0#	77	0.00
31 T	Cyclohexane	0.868	0.788	9.2	65	0.00
32 T	1,1,1-Trichloroethane	0.810	0.885	-9.3	75	0.00
33 S	1,2-Dichloroethane-d4	0.506	0.543	-7.3	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	67	0.00
35 S	Dibromofluoromethane	0.280	0.323	-15.4	75	0.00
36 T	1,1-Dichloropropene	0.469	0.494	-5.3	72	0.00
37 T	Ethyl Acetate	0.284	0.299	-5.3	81	0.00
38 T	Carbon Tetrachloride	0.510	0.559	-9.6	75	0.00
39 T	Methylcyclohexane	0.597	0.581	2.7	66	0.00
40 TM	Benzene	1.343	1.447	-7.7	74	0.00
41 T	Methacrylonitrile	0.154	0.153	0.6	64	-0.01
42 TM	1,2-Dichloroethane	0.436	0.476	-9.2	78	0.00
43 T	Isopropyl Acetate	0.538	0.574	-6.7	81	0.00
44 TM	Trichloroethene	0.367	0.400	-9.0	74	0.00
45 C	1,2-Dichloropropane	0.336	0.374	-11.3#	77	0.00
46 T	Dibromomethane	0.186	0.208	-11.8	80	0.00
47 T	Bromodichloromethane	0.465	0.529	-13.8	78	0.00
48 T	Methyl methacrylate	0.243	0.265	-9.1	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082721\
 Data File : VY005872.D
 Acq On : 27 Aug 2021 10:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 28 03:49:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	86	0.00
50 S	Toluene-d8	1.174	1.276	-8.7	71	0.00
51 T	4-Methyl-2-Pentanone	0.283	0.318	-12.4	85	0.00
52 CM	Toluene	0.856	0.947	-10.6#	75	0.00
53 T	t-1,3-Dichloropropene	0.520	0.574	-10.4	77	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.624	-10.6	76	0.00
55 T	1,1,2-Trichloroethane	0.269	0.307	-14.1	80	0.00
56 T	Ethyl methacrylate	0.402	0.446	-10.9	80	0.00
57 T	1,3-Dichloropropane	0.474	0.535	-12.9	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.192	0.230	-19.8	84	0.00
59 T	2-Hexanone	0.200	0.226	-13.0	85	0.00
60 T	Dibromochloromethane	0.333	0.383	-15.0	82	0.00
61 T	1,2-Dibromoethane	0.255	0.279	-9.4	78	0.00
62 S	4-Bromofluorobenzene	0.389	0.443	-13.9	76	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	70	0.00
64 T	Tetrachloroethene	0.381	0.426	-11.8	80	0.00
65 PM	Chlorobenzene	1.006	1.108	-10.1	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.429	-13.2	79	0.00
67 C	Ethyl Benzene	1.824	1.987	-8.9#	75	0.00
68 T	m/p-Xylenes	0.709	0.787	-11.0	77	0.00
69 T	o-Xylene	0.681	0.753	-10.6	77	0.00
70 T	Styrene	1.146	1.288	-12.4	78	0.00
71 P	Bromoform	0.246	0.279	-13.4	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
73 T	Isopropylbenzene	3.490	3.661	-4.9	76	0.00
74 T	N-acyl acetate	1.051	1.062	-1.0	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.671	0.709	-5.7	85	0.00
76 T	1,2,3-Trichloropropane	0.518	0.480	7.3	74	0.00
77 T	Bromobenzene	0.857	0.902	-5.3	79	0.00
78 T	n-propylbenzene	4.145	4.415	-6.5	78	0.00
79 T	2-Chlorotoluene	2.328	2.470	-6.1	78	0.00
80 T	1,3,5-Trimethylbenzene	2.896	3.073	-6.1	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.257	0.269	-4.7	86	0.00
82 T	4-Chlorotoluene	2.408	2.573	-6.9	79	0.00
83 T	tert-Butylbenzene	2.599	2.753	-5.9	77	0.00
84 T	1,2,4-Trimethylbenzene	2.858	3.060	-7.1	79	0.00
85 T	sec-Butylbenzene	3.749	4.033	-7.6	78	0.00
86 T	p-Isopropyltoluene	3.219	3.489	-8.4	79	0.00
87 T	1,3-Dichlorobenzene	1.673	1.806	-7.9	80	0.00
88 T	1,4-Dichlorobenzene	1.660	1.784	-7.5	81	0.00
89 T	n-Butylbenzene	2.916	3.145	-7.9	79	0.00
90 T	Hexachloroethane	0.579	0.637	-10.0	81	0.00
91 T	1,2-Dichlorobenzene	1.478	1.601	-8.3	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.127	0.128	-0.8	86	0.00
93 T	1,2,4-Trichlorobenzene	1.015	1.103	-8.7	84	0.00
94 T	Hexachlorobutadiene	0.657	0.721	-9.7	81	0.00
95 T	Naphthalene	2.001	2.118	-5.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082721\
 Data File : VY005872.D
 Acq On : 27 Aug 2021 10:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
LabSampleId :
 VSTDCCC050

Quant Time: Aug 28 03:49:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	0.899	0.970	-7.9	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6