

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082724\
 Data File : VY019291.D
 Acq On : 27 Aug 2024 17:15
 Operator : SY/MD
 Sample : P3663-04RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 914-J-SD-0.5-1-081624

Quant Time: Aug 28 01:47:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	17984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	27476	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	20498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	6220	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	7784	43.178	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.360%
35) Dibromofluoromethane	7.616	113	9589	53.375	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.740%
50) Toluene-d8	10.103	98	27252	42.152	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.300%
62) 4-Bromofluorobenzene	12.402	95	6745	30.604	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	61.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082724\
Data File : VY019291.D
Acq On : 27 Aug 2024 17:15
Operator : SY/MD
Sample : P3663-04RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
914-J-SD-0.5-1-081624

Quant Time: Aug 28 01:47:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

