

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082724\
 Data File : VY019306.D
 Acq On : 28 Aug 2024 01:45
 Operator : SY/MD
 Sample : P3717-05
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Aug 28 03:35:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	101813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	177396	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	142358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	64754	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	58167	56.993	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.980%
35) Dibromofluoromethane	7.622	113	57187	49.303	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.600%
50) Toluene-d8	10.103	98	186572	44.697	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.400%
62) 4-Bromofluorobenzene	12.408	95	63469	44.603	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	89.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082724\
Data File : VY019306.D
Acq On : 28 Aug 2024 01:45
Operator : SY/MD
Sample : P3717-05
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Aug 28 03:35:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

