



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082923\
Data File : VY015357.D
Acq On : 29 Aug 2023 16:25
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Title : SW846 8260
Last Update : Thu Aug 31 16:31:54 2023



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	21805	PASS
75	95	30	60	54.4	57685	PASS
95	95	100	100	100.0	106136	PASS
96	95	5	9	6.7	7075	PASS
173	174	0.00	2	1.1	969	PASS
174	95	50	100	82.9	87984	PASS
175	174	5	9	7.6	6706	PASS
176	174	95	101	96.6	84963	PASS
177	176	5	9	7.0	5932	PASS