

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
 Data File : VY005893.D
 Acq On : 30 Aug 2021 15:04
 Operator : SY/MD
 Sample : M3574-01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TPR-12-13-COMP

Quant Time: Aug 31 02:51:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	112122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	158914	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	154740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	75120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	58447	51.497	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.000%
35) Dibromofluoromethane	7.722	113	53870	60.547	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	121.100%
50) Toluene-d8	10.185	98	202707	54.324	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.640%
62) 4-Bromofluorobenzene	12.483	95	64775	52.421	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.840%
Target Compounds						
						Qvalue
17) Carbon Disulfide	4.198	76	3247	1.090	ug/l	# 92
69) o-Xylene	12.032	106	2315	1.098	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
Data File : VY005893.D
Acq On : 30 Aug 2021 15:04
Operator : SY/MD
Sample : M3574-01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TPR-12-13-COMP

Quant Time: Aug 31 02:51:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

