

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
 Data File : VY005894.D
 Acq On : 30 Aug 2021 15:27
 Operator : SY/MD
 Sample : M3577-01
 Misc : 6.13G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 191-0827-B1-0-2

Quant Time: Aug 31 02:51:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	212619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	359436	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	366762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	175648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	129876	60.345	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.680%	
35) Dibromofluoromethane	7.728	113	81058	40.279	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 80.560%	
50) Toluene-d8	10.185	98	456633	54.104	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.200%	
62) 4-Bromofluorobenzene	12.483	95	157963	56.519	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 113.040%	
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	14566	26.843	ug/l	96
44) Trichloroethene	8.947	130	3339	1.267	ug/l	97
64) Tetrachloroethene	10.721	164	61880	22.166	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
Data File : VY005894.D
Acq On : 30 Aug 2021 15:27
Operator : SY/MD
Sample : M3577-01
Misc : 6.13G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
191-0827-B1-0-2

Quant Time: Aug 31 02:51:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

