

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
 Data File : VY005896.D
 Acq On : 30 Aug 2021 16:14
 Operator : SY/MD
 Sample : M3574-01RE
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TPR-12-13-COMP

Quant Time: Aug 31 02:51:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	2719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	4516	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	4821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	1053	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1882	68.379	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 136.760%	
35) Dibromofluoromethane	7.716	113	860	34.014	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 68.020%	
50) Toluene-d8	10.185	98	5882	55.470	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.940%	
62) 4-Bromofluorobenzene	12.483	95	1522	43.343	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 86.680%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
Data File : VY005896.D
Acq On : 30 Aug 2021 16:14
Operator : SY/MD
Sample : M3574-01RE
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TPR-12-13-COMPRE

Quant Time: Aug 31 02:51:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

