

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
 Data File : VY005898.D
 Acq On : 30 Aug 2021 16:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 31 02:51:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	64	0.00
2 T	Dichlorodifluoromethane	0.411	0.357	13.1	62	0.00
3 P	Chloromethane	0.491	0.438	10.8	59	0.00
4 C	Vinyl Chloride	0.588	0.576	2.0#	64	0.00
5 T	Bromomethane	0.387	0.420	-8.5	72	0.00
6 T	Chloroethane	0.355	0.402	-13.2	69	0.00
7 T	Trichlorofluoromethane	0.787	0.855	-8.6	68	0.00
8 T	Diethyl Ether	0.275	0.336	-22.2	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.427	0.495	-15.9	77	0.00
10 T	Methyl Iodide	0.532	0.561	-5.5	69	0.00
11 T	Tert butyl alcohol	0.050	0.082	-64.0#	132	0.00
12 CM	1,1-Dichloroethene	0.431	0.485	-12.5#	75	0.00
13 T	Acrolein	0.030	0.032	-6.7	70	0.00
14 T	Allyl chloride	0.755	0.881	-16.7	77	0.00
15 T	Acrylonitrile	0.141	0.191	-35.5#	99	0.00
16 T	Acetone	0.128	0.156	-21.9	84	0.00
17 T	Carbon Disulfide	1.328	1.249	5.9	63	0.00
18 T	Methyl Acetate	0.334	0.440	-31.7#	103	0.00
19 T	Methyl tert-butyl Ether	1.335	1.690	-26.6#	87	0.00
20 T	Methylene Chloride	0.641	0.656	-2.3	83	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.548	-13.7	76	0.00
22 T	Diisopropyl ether	1.557	1.944	-24.9	83	0.00
23 T	Vinyl Acetate	1.036	1.306	-26.1#	86	0.00
24 P	1,1-Dichloroethane	0.863	1.079	-25.0#	82	0.00
25 T	2-Butanone	0.189	0.211	-11.6	81	0.00
26 T	2,2-Dichloropropane	0.788	0.795	-0.9	66	0.00
27 T	cis-1,2-Dichloroethene	0.548	0.580	-5.8	70	0.00
28 T	Bromochloromethane	0.344	0.307	10.8	65	0.00
29 T	Tetrahydrofuran	0.125	0.140	-12.0	86	0.00
30 C	Chloroform	0.881	0.957	-8.6#	72	0.00
31 T	Cyclohexane	0.868	0.737	15.1	59	0.00
32 T	1,1,1-Trichloroethane	0.810	0.848	-4.7	69	0.00
33 S	1,2-Dichloroethane-d4	0.506	0.539	-6.5	70	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	65	0.00
35 S	Dibromofluoromethane	0.280	0.308	-10.0	69	0.00
36 T	1,1-Dichloropropene	0.469	0.461	1.7	65	0.00
37 T	Ethyl Acetate	0.284	0.319	-12.3	84	0.00
38 T	Carbon Tetrachloride	0.510	0.528	-3.5	69	0.00
39 T	Methylcyclohexane	0.597	0.547	8.4	60	0.00
40 TM	Benzene	1.343	1.376	-2.5	68	0.00
41 T	Methacrylonitrile	0.154	0.179	-16.2	73	0.00
42 TM	1,2-Dichloroethane	0.436	0.466	-6.9	74	0.00
43 T	Isopropyl Acetate	0.538	0.587	-9.1	80	0.00
44 TM	Trichloroethene	0.367	0.382	-4.1	69	0.00
45 C	1,2-Dichloropropane	0.336	0.352	-4.8#	70	0.00
46 T	Dibromomethane	0.186	0.204	-9.7	76	0.00
47 T	Bromodichloromethane	0.465	0.509	-9.5	73	0.00
48 T	Methyl methacrylate	0.243	0.271	-11.5	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
 Data File : VY005898.D
 Acq On : 30 Aug 2021 16:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 31 02:51:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	90	0.00
50 S	Toluene-d8	1.174	1.200	-2.2	65	0.00
51 T	4-Methyl-2-Pentanone	0.283	0.332	-17.3	86	0.00
52 CM	Toluene	0.856	0.899	-5.0#	69	0.00
53 T	t-1,3-Dichloropropene	0.520	0.557	-7.1	72	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.597	-5.9	70	0.00
55 T	1,1,2-Trichloroethane	0.269	0.302	-12.3	77	0.00
56 T	Ethyl methacrylate	0.402	0.448	-11.4	78	0.00
57 T	1,3-Dichloropropane	0.474	0.520	-9.7	75	0.00
58 T	2-Chloroethyl Vinyl ether	0.192	0.238	-24.0	84	0.00
59 T	2-Hexanone	0.200	0.234	-17.0	86	0.00
60 T	Dibromochloromethane	0.333	0.377	-13.2	78	0.00
61 T	1,2-Dibromoethane	0.255	0.279	-9.4	76	0.00
62 S	4-Bromofluorobenzene	0.389	0.428	-10.0	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	68	0.00
64 T	Tetrachloroethene	0.381	0.399	-4.7	73	0.00
65 PM	Chlorobenzene	1.006	1.051	-4.5	72	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.416	-9.8	76	0.00
67 C	Ethyl Benzene	1.824	1.872	-2.6#	70	0.00
68 T	m/p-Xylenes	0.709	0.727	-2.5	69	0.00
69 T	o-Xylene	0.681	0.709	-4.1	71	0.00
70 T	Styrene	1.146	1.235	-7.8	74	0.00
71 P	Bromoform	0.246	0.278	-13.0	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
73 T	Isopropylbenzene	3.490	3.445	1.3	70	0.00
74 T	N-amyl acetate	1.051	1.071	-1.9	78	0.00
75 P	1,1,2,2-Tetrachloroethane	0.671	0.722	-7.6	84	0.00
76 T	1,2,3-Trichloropropane	0.518	0.485	6.4	73	0.00
77 T	Bromobenzene	0.857	0.882	-2.9	75	0.00
78 T	n-propylbenzene	4.145	4.144	0.0	71	0.00
79 T	2-Chlorotoluene	2.328	2.330	-0.1	72	0.00
80 T	1,3,5-Trimethylbenzene	2.896	2.921	-0.9	72	0.00
81 T	trans-1,4-Dichloro-2-butene	0.257	0.266	-3.5	82	0.00
82 T	4-Chlorotoluene	2.408	2.395	0.5	72	0.00
83 T	tert-Butylbenzene	2.599	2.605	-0.2	72	0.00
84 T	1,2,4-Trimethylbenzene	2.858	2.898	-1.4	73	0.00
85 T	sec-Butylbenzene	3.749	3.767	-0.5	71	0.00
86 T	p-Isopropyltoluene	3.219	3.260	-1.3	72	0.00
87 T	1,3-Dichlorobenzene	1.673	1.710	-2.2	74	0.00
88 T	1,4-Dichlorobenzene	1.660	1.691	-1.9	75	0.00
89 T	n-Butylbenzene	2.916	2.930	-0.5	72	0.00
90 T	Hexachloroethane	0.579	0.599	-3.5	74	0.00
91 T	1,2-Dichlorobenzene	1.478	1.548	-4.7	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.127	0.138	-8.7	90	0.00
93 T	1,2,4-Trichlorobenzene	1.015	1.055	-3.9	78	0.00
94 T	Hexachlorobutadiene	0.657	0.668	-1.7	73	0.00
95 T	Naphthalene	2.001	2.162	-8.0	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083021\
Data File : VY005898.D
Acq On : 30 Aug 2021 16:59
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 31 02:51:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.936	-4.1	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6