

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083023\
 Data File : VY015371.D
 Acq On : 30 Aug 2023 12:41
 Operator : SY/MD
 Sample : 04203-01
 Misc : 5.42g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-1(12-12.5)

Quant Time: Aug 31 00:59:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 29 22:47:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	136098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	224876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	201298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	96913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	87581	60.030	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.060%	
35) Dibromofluoromethane	7.716	113	70338	54.305	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.620%	
50) Toluene-d8	10.179	98	262304	55.922	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 111.840%	
62) 4-Bromofluorobenzene	12.477	95	88178	50.639	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.280%	
Target Compounds						
20) Methylene Chloride	4.710	84	32804	11.771	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083023\
Data File : VY015371.D
Acq On : 30 Aug 2023 12:41
Operator : SY/MD
Sample : 04203-01
Misc : 5.42g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-1(12-12.5)

Quant Time: Aug 31 00:59:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 29 22:47:33 2023
Response via : Initial Calibration

