



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083023\
Data File : VY015366.D
Acq On : 30 Aug 2023 08:47
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Title : SW846 8260
Last Update : Thu Aug 31 20:01:47 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.7	20133	PASS
75	95	30	60	54.7	55893	PASS
95	95	100	100	100.0	102272	PASS
96	95	5	9	6.4	6594	PASS
173	174	0.00	2	1.1	974	PASS
174	95	50	100	85.3	87219	PASS
175	174	5	9	7.4	6447	PASS
176	174	95	101	96.2	83896	PASS
177	176	5	9	7.3	6107	PASS