

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083121\
 Data File : VY005912.D
 Acq On : 31 Aug 2021 15:10
 Operator : SY/MD
 Sample : M3602-09
 Misc : 4.83G/5ML/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1COMP

Quant Time: Sep 01 05:37:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	199038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	338126	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	358094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	184866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	124923	62.282	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.560%	
35) Dibromofluoromethane	7.728	113	104329	54.942	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.880%	
50) Toluene-d8	10.185	98	436776	54.807	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.620%	
62) 4-Bromofluorobenzene	12.483	95	157816	58.975	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 117.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083121\
Data File : VY005912.D
Acq On : 31 Aug 2021 15:10
Operator : SY/MD
Sample : M3602-09
Misc : 4.83G/5ML/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1COMP

Quant Time: Sep 01 05:37:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

