

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083122\
 Data File : VY010252.D
 Acq On : 31 Aug 2022 17:51
 Operator : KP/MD
 Sample : N4424-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-219

Quant Time: Aug 31 23:42:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 02:23:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	139263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	227515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	216240	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	102835	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	72801	62.013	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.020%	
35) Dibromofluoromethane	7.709	113	72735	55.526	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.060%	
50) Toluene-d8	10.178	98	267017	54.845	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.680%	
62) 4-Bromofluorobenzene	12.483	95	86670	47.814	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 95.620%	
Target Compounds						
20) Methylene Chloride	4.704	84	9208	6.343	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083122\
Data File : VY010252.D
Acq On : 31 Aug 2022 17:51
Operator : KP/MD
Sample : N4424-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-219

Quant Time: Aug 31 23:42:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 02:23:17 2022
Response via : Initial Calibration

