

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083122\
 Data File : VY010232.D
 Acq On : 31 Aug 2022 09:37
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY083122

Quant Time: Aug 31 14:43:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 02:23:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.365	0.349	4.4	104	0.00
3 P	Chloromethane	0.371	0.341	8.1	98	0.00
4 C	Vinyl Chloride	0.417	0.394	5.5#	95	0.00
5 T	Bromomethane	0.338	0.307	9.2	100	0.00
6 T	Chloroethane	0.283	0.280	1.1	100	0.01
7 T	Trichlorofluoromethane	0.782	0.755	3.5	97	0.01
8 T	Diethyl Ether	0.255	0.245	3.9	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.444	0.451	-1.6	104	0.01
10 T	Methyl Iodide	0.578	0.622	-7.6	101	0.00
11 T	Tert butyl alcohol	0.042	0.033	21.4	86	0.00
12 CM	1,1-Dichloroethene	0.447	0.436	2.5#	101	0.00
13 T	Acrolein	0.055	0.053	3.6	92	0.00
14 T	Allyl chloride	0.610	0.599	1.8	102	0.00
15 T	Acrylonitrile	0.116	0.109	6.0	97	0.00
16 T	Acetone	0.089	0.100	-12.4	117	0.00
17 T	Carbon Disulfide	1.344	1.262	6.1	98	0.00
18 T	Methyl Acetate	0.291	0.265	8.9	95	0.00
19 T	Methyl tert-butyl Ether	1.159	1.135	2.1	98	0.00
20 T	Methylene Chloride	0.521	0.477	8.4	99	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.488	0.6	101	0.00
22 T	Diisopropyl ether	1.227	1.222	0.4	101	0.00
23 T	Vinyl Acetate	0.786	0.764	2.8	97	0.00
24 P	1,1-Dichloroethane	0.769	0.762	0.9	102	0.00
25 T	2-Butanone	0.152	0.135	11.2	99	0.00
26 T	2,2-Dichloropropane	0.719	0.746	-3.8	109	0.00
27 T	cis-1,2-Dichloroethene	0.558	0.544	2.5	101	0.00
28 T	Bromochloromethane	0.260	0.269	-3.5	102	0.00
29 T	Tetrahydrofuran	0.101	0.086	14.9	91	0.00
30 C	Chloroform	0.871	0.854	2.0#	103	0.00
31 T	Cyclohexane	0.752	0.667	11.3	97	0.00
32 T	1,1,1-Trichloroethane	0.798	0.799	-0.1	102	0.00
33 S	1,2-Dichloroethane-d4	0.381	0.405	-6.3	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.263	0.280	-6.5	92	0.00
36 T	1,1-Dichloropropene	0.436	0.444	-1.8	101	0.00
37 T	Ethyl Acetate	0.230	0.212	7.8	94	0.00
38 T	Carbon Tetrachloride	0.500	0.527	-5.4	102	0.00
39 T	Methylcyclohexane	0.559	0.549	1.8	99	0.00
40 TM	Benzene	1.288	1.280	0.6	100	0.00
41 T	Methacrylonitrile	0.130	0.132	-1.5	104	0.02
42 TM	1,2-Dichloroethane	0.381	0.380	0.3	98	0.00
43 T	Isopropyl Acetate	0.440	0.388	11.8	92	0.00
44 TM	Trichloroethene	0.397	0.399	-0.5	100	0.00
45 C	1,2-Dichloropropane	0.311	0.289	7.1#	100	0.00
46 T	Dibromomethane	0.192	0.190	1.0	100	0.00
47 T	Bromodichloromethane	0.445	0.528	-18.7	118	0.00
48 T	Methyl methacrylate	0.196	0.182	7.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083122\
 Data File : VY010232.D
 Acq On : 31 Aug 2022 09:37
 Operator : KP/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY083122

Quant Time: Aug 31 14:43:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 02:23:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	103	0.00
50 S	Toluene-d8	0.925	1.268	-37.1#	109	0.00
51 T	4-Methyl-2-Pentanone	0.235	0.269	-14.5	119	0.00
52 CM	Toluene	0.840	0.977	-16.3#	116	0.00
53 T	t-1,3-Dichloropropene	0.487	0.560	-15.0	121	0.00
54 T	cis-1,3-Dichloropropene	0.523	0.632	-20.8	123	0.00
55 T	1,1,2-Trichloroethane	0.293	0.309	-5.5	114	0.00
56 T	Ethyl methacrylate	0.380	0.420	-10.5	116	0.00
57 T	1,3-Dichloropropane	0.482	0.531	-10.2	120	0.00
58 T	2-Chloroethyl Vinyl ether	0.117	0.104	11.1	120	0.00
59 T	2-Hexanone	0.179	0.204	-14.0	128	0.00
60 T	Dibromochloromethane	0.367	0.401	-9.3	112	0.00
61 T	1,2-Dibromoethane	0.289	0.306	-5.9	114	0.00
62 S	4-Bromofluorobenzene	0.398	0.433	-8.8	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	117	0.00
64 T	Tetrachloroethene	0.432	0.423	2.1	111	0.00
65 PM	Chlorobenzene	1.030	1.024	0.6	117	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.388	0.3	112	0.00
67 C	Ethyl Benzene	1.758	1.765	-0.4#	117	0.00
68 T	m/p-Xylenes	0.701	0.698	0.4	115	0.00
69 T	o-Xylene	0.653	0.661	-1.2	114	0.00
70 T	Styrene	1.094	1.126	-2.9	115	0.00
71 P	Bromoform	0.257	0.254	1.2	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.380	3.533	-4.5	116	0.00
74 T	N-amyl acetate	0.880	0.863	1.9	121	0.00
75 P	1,1,2,2-Tetrachloroethane	0.685	0.655	4.4	117	0.00
76 T	1,2,3-Trichloropropane	0.492	0.511	-3.9	125	0.00
77 T	Bromobenzene	0.864	0.874	-1.2	110	0.00
78 T	n-propylbenzene	4.010	4.253	-6.1	120	0.00
79 T	2-Chlorotoluene	2.313	2.432	-5.1	118	0.00
80 T	1,3,5-Trimethylbenzene	2.872	2.940	-2.4	113	0.00
81 T	trans-1,4-Dichloro-2-butene	0.236	0.242	-2.5	122	0.00
82 T	4-Chlorotoluene	2.427	2.437	-0.4	114	0.00
83 T	tert-Butylbenzene	2.540	2.545	-0.2	110	0.00
84 T	1,2,4-Trimethylbenzene	2.820	2.893	-2.6	114	0.00
85 T	sec-Butylbenzene	3.736	3.734	0.1	113	0.00
86 T	p-Isopropyltoluene	3.096	3.296	-6.5	116	0.00
87 T	1,3-Dichlorobenzene	1.707	1.760	-3.1	117	0.00
88 T	1,4-Dichlorobenzene	1.720	1.701	1.1	112	0.00
89 T	n-Butylbenzene	2.866	2.985	-4.2	113	0.00
90 T	Hexachloroethane	0.624	0.626	-0.3	102	0.00
91 T	1,2-Dichlorobenzene	1.544	1.566	-1.4	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.119	0.109	8.4	93	0.00
93 T	1,2,4-Trichlorobenzene	1.013	0.978	3.5	99	0.00
94 T	Hexachlorobutadiene	0.600	0.567	5.5	100	0.00
95 T	Naphthalene	1.997	1.741	12.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083122\
Data File : VY010232.D
Acq On : 31 Aug 2022 09:37
Operator : KP/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY083122

Quant Time: Aug 31 14:43:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 02:23:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.799	11.1	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6