

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
 Data File : VY015397.D
 Acq On : 31 Aug 2023 12:41
 Operator : SY/MD
 Sample : 04249-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-02-083023

Quant Time: Sep 01 00:45:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	48413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	78640	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	72238	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	33560	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	32303	62.243	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.480%
35) Dibromofluoromethane	7.716	113	26398	58.280	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.560%
50) Toluene-d8	10.179	98	93188	56.812	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	113.620%
62) 4-Bromofluorobenzene	12.483	95	30040	49.332	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.660%
Target Compounds						
16) Acetone	3.942	43	3902	18.013	ug/l	99
20) Methylene Chloride	4.716	84	9292	8.007	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
Data File : VY015397.D
Acq On : 31 Aug 2023 12:41
Operator : SY/MD
Sample : 04249-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-02-083023

Quant Time: Sep 01 00:45:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

