

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
 Data File : VY015399.D
 Acq On : 31 Aug 2023 13:28
 Operator : SY/MD
 Sample : 04202-01
 Misc : 6.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C-1

Quant Time: Sep 01 00:46:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	118191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	202137	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	182684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	86876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	79353	62.631	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.260%	
35) Dibromofluoromethane	7.710	113	65206	56.006	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.020%	
50) Toluene-d8	10.179	98	237992	56.446	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.900%	
62) 4-Bromofluorobenzene	12.477	95	77495	49.511	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.020%	
Target Compounds						
20) Methylene Chloride	4.704	84	26846	10.703	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
Data File : VY015399.D
Acq On : 31 Aug 2023 13:28
Operator : SY/MD
Sample : 04202-01
Misc : 6.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C-1

Quant Time: Sep 01 00:46:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

