

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
 Data File : VY015391.D
 Acq On : 31 Aug 2023 09:41
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/01/2023
 Supervised By : Mahesh Dadoda 09/01/2023

Quant Time: Sep 01 00:42:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	159774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	269421	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	250765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	88290	51.549	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 103.100%			
35) Dibromofluoromethane	7.716	113	80684	51.994	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 103.980%			
50) Toluene-d8	10.179	98	274697	48.881	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 97.760%			
62) 4-Bromofluorobenzene	12.483	95	104357	50.022	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 100.040%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	62697	45.143	ug/l	100
3) Chloromethane	2.113	50	95111	51.565	ug/l	100
4) Vinyl Chloride	2.247	62	102270	47.636	ug/l	99
5) Bromomethane	2.643	94	76858	50.313	ug/l	100
6) Chloroethane	2.784	64	71261	50.499	ug/l	96
7) Trichlorofluoromethane	3.119	101	138928	45.748	ug/l	100
8) Diethyl Ether	3.527	74	54574	52.584	ug/l	93
9) 1,1,2-Trichlorotrifluo...	3.893	101	77011	45.754	ug/l	95
10) Methyl Iodide	4.088	142	97295	49.224	ug/l	98
11) Tert butyl alcohol	4.972	59	104512	281.486	ug/l #	88
12) 1,1-Dichloroethene	3.869	96	76242	48.003	ug/l	92
13) Acrolein	3.735	56	153214	318.640	ug/l	98
14) Allyl chloride	4.479	41	131330	51.934	ug/l	95
15) Acrylonitrile	5.167	53	190124	291.128	ug/l	99
16) Acetone	3.960	43	262494	367.178	ug/l	94
17) Carbon Disulfide	4.192	76	230347	48.512	ug/l	100
18) Methyl Acetate	4.485	43	147792	57.076	ug/l	94
19) Methyl tert-butyl Ether	5.216	73	272620	52.812	ug/l	100
20) Methylene Chloride	4.716	84	101001	42.930	ug/l	92
21) trans-1,2-Dichloroethene	5.216	96	91649	50.111	ug/l	98
22) Diisopropyl ether	6.118	45	288392	52.145	ug/l	93
23) Vinyl Acetate	6.058	43	978882	280.328	ug/l	95
24) 1,1-Dichloroethane	6.015	63	159976	50.038	ug/l	96
25) 2-Butanone	6.990	43	340261	335.577	ug/l	98
26) 2,2-Dichloropropane	6.984	77	140734	51.827	ug/l	99
27) cis-1,2-Dichloroethene	6.984	96	106832	51.490	ug/l	95
28) Bromochloromethane	7.332	49	50889	51.905	ug/l	98
29) Tetrahydrofuran	7.350	42	188085	295.388	ug/l	94
30) Chloroform	7.502	83	170846	50.833	ug/l	97
31) Cyclohexane	7.783	56	134249	44.126	ug/l	91
32) 1,1,1-Trichloroethane	7.704	97	145275	48.393	ug/l	99
36) 1,1-Dichloropropene	7.917	75	125268	45.973	ug/l	99
37) Ethyl Acetate	7.076	43	120872	57.517	ug/l	97
38) Carbon Tetrachloride	7.899	117	129282	45.939	ug/l	99
39) Methylcyclohexane	9.185	83	151106	44.049	ug/l	94
40) Benzene	8.161	78	369740	48.636	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
 Data File : VY015391.D
 Acq On : 31 Aug 2023 09:41
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/01/2023
 Supervised By : Mahesh Dadoda 09/01/2023

Quant Time: Sep 01 00:42:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.307	41	55387m	56.203	ug/l	
42) 1,2-Dichloroethane	8.234	62	124199	48.994	ug/l	95
43) Isopropyl Acetate	8.271	43	198248	54.554	ug/l	98
44) Trichloroethene	8.941	130	103491	48.115	ug/l	97
45) 1,2-Dichloropropane	9.215	63	96429	50.317	ug/l	99
46) Dibromomethane	9.307	93	64410	50.597	ug/l	96
47) Bromodichloromethane	9.496	83	136940	50.302	ug/l	96
48) Methyl methacrylate	9.295	41	89530	53.697	ug/l	92
49) 1,4-Dioxane	9.307	88	26721	1130.071	ug/l #	49
51) 4-Methyl-2-Pentanone	10.069	43	613247	290.344	ug/l	94
52) Toluene	10.240	92	235920	48.419	ug/l	96
53) t-1,3-Dichloropropene	10.465	75	151043	52.206	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	161427	51.292	ug/l	92
55) 1,1,2-Trichloroethane	10.642	97	88321	51.491	ug/l	96
56) Ethyl methacrylate	10.508	69	136446	53.680	ug/l	90
57) 1,3-Dichloropropane	10.788	76	151310	51.246	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.782	63	248461	273.857	ug/l	95
59) 2-Hexanone	10.831	43	500411	315.748	ug/l	93
60) Dibromochloromethane	10.983	129	104842	52.222	ug/l	100
61) 1,2-Dibromoethane	11.087	107	91490	52.640	ug/l	99
64) Tetrachloroethene	10.721	164	108177	47.527	ug/l	97
65) Chlorobenzene	11.514	112	258230	47.788	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.587	131	95888	48.481	ug/l	98
67) Ethyl Benzene	11.593	91	457577	47.158	ug/l	98
68) m/p-Xylenes	11.703	106	356228	94.715	ug/l	95
69) o-Xylene	12.032	106	171348	47.671	ug/l	97
70) Styrene	12.044	104	296466	49.054	ug/l	98
71) Bromoform	12.209	173	74447	53.960	ug/l #	98
73) Isopropylbenzene	12.331	105	447222	46.111	ug/l	100
74) N-amyl acetate	12.142	43	179062	54.953	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.581	83	123710	53.432	ug/l	100
76) 1,2,3-Trichloropropane	12.636	75	96694m	56.470	ug/l	
77) Bromobenzene	12.611	156	110952	48.301	ug/l	95
78) n-propylbenzene	12.672	91	552356	47.066	ug/l	100
79) 2-Chlorotoluene	12.758	91	313914	47.689	ug/l	99
80) 1,3,5-Trimethylbenzene	12.812	105	380073	46.851	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.380	75	43585	56.149	ug/l	97
82) 4-Chlorotoluene	12.855	91	328141	47.966	ug/l	100
83) tert-Butylbenzene	13.075	119	323925	45.588	ug/l	99
84) 1,2,4-Trimethylbenzene	13.117	105	384905	47.787	ug/l	99
85) sec-Butylbenzene	13.251	105	489326	46.171	ug/l	99
86) p-Isopropyltoluene	13.367	119	413016	46.498	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	217229	47.901	ug/l	99
88) 1,4-Dichlorobenzene	13.446	146	219520	47.887	ug/l	100
89) n-Butylbenzene	13.696	91	393681	47.090	ug/l	97
90) Hexachloroethane	13.959	117	77100	48.533	ug/l	88
91) 1,2-Dichlorobenzene	13.739	146	202893	49.025	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	25522	55.724	ug/l	90
93) 1,2,4-Trichlorobenzene	15.007	180	130322	50.142	ug/l	99
94) Hexachlorobutadiene	15.111	225	65458	46.187	ug/l	98
95) Naphthalene	15.239	128	355843	55.271	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	120400	50.787	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
Data File : VY015391.D
Acq On : 31 Aug 2023 09:41
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023

Quant Time: Sep 01 00:42:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083123\
 Data File : VY015391.D
 Acq On : 31 Aug 2023 09:41
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/soil
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations APPROVED

Reviewed By : John Carlone 09/01/2023
 Supervised By : Mahesh Dadoda 09/01/2023


Quant Time: Sep 01 00:42:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 159774
Ion Ratio Lower Upper
168 100
99 56.8 43.4 65.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#2
Dichlorodifluoromethane
Concen: 45.143 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 85 Resp: 62697
Ion Ratio Lower Upper
85 100
87 31.0 15.6 46.8





#3
Chloromethane
Concen: 51.565 ug/l
RT: 2.113 min Scan# 41
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 9511
Ion Ratio Lower Upper
50 100
52 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#4
Vinyl Chloride
Concen: 47.636 ug/l
RT: 2.247 min Scan# 70
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 62 Resp: 102270
Ion Ratio Lower Upper
62 100
64 30.9 24.3 36.5





#5
Bromomethane
Concen: 50.313 ug/l
RT: 2.643 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 7685
Ion Ratio Lower Upper
94 100
96 94.4 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#6
Chloroethane
Concen: 50.499 ug/l
RT: 2.784 min Scan# 158
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 64 Resp: 71261
Ion Ratio Lower Upper
64 100
66 32.9 24.4 36.6





#7
Trichlorofluoromethane
Concen: 45.748 ug/l
RT: 3.119 min Scan# 21
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 138928
Ion Ratio Lower Upper
101 100
103 64.6 51.4 77.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#8
Diethyl Ether
Concen: 52.584 ug/l
RT: 3.527 min Scan# 280
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 74 Resp: 54574
Ion Ratio Lower Upper
74 100
45 95.9 51.5 154.5





#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.754 ug/l
RT: 3.893 min Scan# 340
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 7701
Ion Ratio Lower Upper
101 100
85 45.5 36.6 55.0
151 76.9 67.0 100.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#10
Methyl Iodide
Concen: 49.224 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Tgt Ion:142 Resp: 97295
Ion Ratio Lower Upper
142 100
127 45.4 37.6 56.4
141 14.6 11.8 17.8









#11
Tert butyl alcohol
Concen: 281.486 ug/l
RT: 4.972 min Scan# 515
Delta R.T. 0.018 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 59 Resp: 104512
Ion Ratio Lower Upper
59 100
57 7.7 3.0 4.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#12
1,1-Dichloroethene
Concen: 48.003 ug/l
RT: 3.869 min Scan# 336
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 96 Resp: 76242
Ion Ratio Lower Upper
96 100
61 154.8 136.4 204.6
98 61.5 49.4 74.2





#13
Acrolein
Concen: 318.640 ug/l
RT: 3.735 min Scan# 312
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 153214
Ion Ratio Lower Upper
56 100
55 69.8 57.0 85.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#14
Allyl chloride
Concen: 51.934 ug/l
RT: 4.479 min Scan# 436
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion: 41 Resp: 131330
Ion Ratio Lower Upper
41 100
39 68.5 58.0 87.0
76 37.4 27.4 41.2









#15
Acrylonitrile
Concen: 291.128 ug/l
RT: 5.167 min Scan# 547
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 53 Resp: 190124
Ion Ratio Lower Upper
53 100
52 82.1 66.9 100.3
51 36.6 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#16
Acetone
Concen: 367.178 ug/l
RT: 3.960 min Scan# 351
Delta R.T. 0.018 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 43 Resp: 262494
Ion Ratio Lower Upper
43 100
58 31.6 22.7 34.1





#17
Carbon Disulfide
Concen: 48.512 ug/l
RT: 4.192 min Scan# 31
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 23034
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#18
Methyl Acetate
Concen: 57.076 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 43 Resp: 147792
Ion Ratio Lower Upper
43 100
74 24.4 17.4 26.0





#19
Methyl tert-butyl Ether
Concen: 52.812 ug/l
RT: 5.216 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 73 Resp: 272620
Ion Ratio Lower Upper
73 100
57 21.8 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#20
Methylene Chloride
Concen: 42.930 ug/l
RT: 4.716 min Scan# 475
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 84 Resp: 101001
Ion Ratio Lower Upper
84 100
49 121.2 106.9 160.3
51 34.7 33.2 49.8
86 62.3 51.0 76.6





#21
trans-1,2-Dichloroethene
Concen: 50.111 ug/l
RT: 5.216 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 96 Resp: 91649
Ion Ratio Lower Upper
96 100
61 135.1 110.2 165.4
98 66.3 51.8 77.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#22
Diisopropyl ether
Concen: 52.145 ug/l
RT: 6.118 min Scan# 705
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion: 45 Resp: 288392
Ion Ratio Lower Upper
45 100
43 58.7 52.2 78.4
87 28.5 21.0 31.6
59 11.5 8.9 13.3









#23
 Vinyl Acetate
 Concen: 280.328 ug/l
 RT: 6.058 min Scan# 694
 Delta R.T. 0.006 min
 Lab File: VY015391.D
 Acq: 31 Aug 2023 09:41

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Tgt Ion: 43 Resp: 97888
 Ion Ratio Lower Upper
 43 100
 86 11.2 7.5 11.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 09/01/2023
 Supervised By :Mahesh Dadoda 09/01/2023



#24
 1,1-Dichloroethane
 Concen: 50.038 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. 0.012 min
 Lab File: VY015391.D
 Acq: 31 Aug 2023 09:41


Tgt Ion: 63 Resp: 159976
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.1 9.4
 100 5.1 1.8 5.5





#25
2-Butanone
Concen: 335.577 ug/l
RT: 6.990 min Scan# 846
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 34026
Ion Ratio Lower Upper
43 100
72 26.5 20.5 30.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#26
2,2-Dichloropropane
Concen: 51.827 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion: 77 Resp: 140734
Ion Ratio Lower Upper
77 100
97 22.3 10.9 32.9






#27
cis-1,2-Dichloroethene
Concen: 51.490 ug/l
RT: 6.984 min Scan# 846
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050






Tgt Ion: 96 Resp: 10683
Ion Ratio Lower Upper
96 100
61 141.4 0.0 301.4
98 64.4 0.0 130.4


Manual Integrations

APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#28
Bromochloromethane
Concen: 51.905 ug/l
RT: 7.332 min Scan# 904
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 49 Resp: 50889
Ion Ratio Lower Upper
49 100
129 2.0 0.0 4.6
130 78.0 60.7 91.1





#29
Tetrahydrofuran
Concen: 295.388 ug/l
RT: 7.350 min Scan# 907
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 42 Resp: 188085
Ion Ratio Lower Upper
42 100
72 45.1 32.8 49.2
71 41.8 31.0 46.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#30
Chloroform
Concen: 50.833 ug/l
RT: 7.502 min Scan# 932
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 83 Resp: 170846
Ion Ratio Lower Upper
83 100
85 64.2 49.7 74.5





#31
Cyclohexane
Concen: 44.126 ug/l
RT: 7.783 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 134249
Ion Ratio Lower Upper
56 100
69 32.8 27.8 41.6
84 92.7 66.4 99.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#32
1,1,1-Trichloroethane
Concen: 48.393 ug/l
RT: 7.704 min Scan# 965
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 97 Resp: 145275
Ion Ratio Lower Upper
97 100
99 64.6 50.8 76.2
61 42.9 35.1 52.7





#33
1,2-Dichloroethane-d4
Concen: 51.549 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 65 Resp: 88290
Ion Ratio Lower Upper
65 100
67 52.5 0.0 101.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion:114 Resp: 269421
Ion Ratio Lower Upper
114 100
63 19.4 0.0 42.4
88 14.8 0.0 30.6





#35
Dibromofluoromethane
Concen: 51.994 ug/l
RT: 7.716 min Scan# 90
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:113 Resp: 80684
Ion Ratio Lower Upper
113 100
111 101.5 81.3 121.9
192 18.7 16.9 25.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#36
1,1-Dichloropropene
Concen: 45.973 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 75 Resp: 125268
Ion Ratio Lower Upper
75 100
110 34.9 17.8 53.5
77 31.2 24.4 36.6





#37
Ethyl Acetate
Concen: 57.517 ug/l
RT: 7.076 min Scan# 861
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 120872
Ion Ratio Lower Upper
43 100
61 13.8 10.1 15.1
70 9.9 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#38
Carbon Tetrachloride
Concen: 45.939 ug/l
RT: 7.899 min Scan# 997
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Tgt Ion:117 Resp: 129282
Ion Ratio Lower Upper
117 100
119 96.4 76.4 114.6
121 31.4 25.8 38.6





#39

Methylcyclohexane

Concen: 44.049 ug/l

RT: 9.185 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY015391.D

Acq: 31 Aug 2023 09:41

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 151100

Ion Ratio Lower Upper

83 100

55 72.6 63.0 94.4





98 49.3 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/01/2023

Supervised By :Mahesh Dadoda 09/01/2023



#40

Benzene

Concen: 48.636 ug/l

RT: 8.161 min Scan# 1040

Delta R.T. 0.012 min

Lab File: VY015391.D




Acq: 31 Aug 2023 09:41


Tgt Ion: 78 Resp: 369740

Ion Ratio Lower Upper

78 100

77 24.0 19.0 28.6






#41
Methacrylonitrile
Concen: 56.203 ug/l m
RT: 7.307 min Scan# 900
Delta R.T. -0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050






Tgt Ion: 41 Resp: 5538
Ion Ratio Lower Upper
41 100
39 81.2 52.1 78.1
67 0.0 66.2 99.2
52 0.0 29.9 44.9

Manual Integrations




APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#42
1,2-Dichloroethane
Concen: 48.994 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 62 Resp: 124199
Ion Ratio Lower Upper
62 100
98 9.8 0.0 15.8





#43
Isopropyl Acetate
Concen: 54.554 ug/l
RT: 8.271 min Scan# 1058
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 198248
Ion Ratio Lower Upper
43 100
61 27.0 21.1 31.7
87 13.6 9.8 14.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#44
Trichloroethene
Concen: 48.115 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion:130 Resp: 103491
Ion Ratio Lower Upper
130 100
95 94.9 0.0 196.0





#45
1,2-Dichloropropane
Concen: 50.317 ug/l
RT: 9.215 min Scan# 1212
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 63 Resp: 96429
Ion Ratio Lower Upper
63 100
65 30.5 23.8 35.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#46
Dibromomethane
Concen: 50.597 ug/l
RT: 9.307 min Scan# 1228
Delta R.T. 0.012 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion: 93 Resp: 64410
Ion Ratio Lower Upper
93 100
95 83.4 66.6 100.0
174 98.8 84.8 127.2









#47
Bromodichloromethane
Concen: 50.302 ug/l
RT: 9.496 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 136940
Ion Ratio Lower Upper
83 100
85 65.6 50.1 75.1
127 8.0 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#48
Methyl methacrylate
Concen: 53.697 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 41 Resp: 89530
Ion Ratio Lower Upper
41 100
69 88.3 64.5 96.7
39 57.2 50.8 76.2





#49
1,4-Dioxane
Concen: 1130.071 ug/l
RT: 9.307 min Scan# 1228
Delta R.T. 0.018 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 88 Resp: 2672
Ion Ratio Lower Upper
88 100
43 22.5 0.0 0.0
58 69.8 30.9 46.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#50
Toluene-d8
Concen: 48.881 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 98 Resp: 274697
Ion Ratio Lower Upper
98 100
100 64.9 50.1 75.1





#51
4-Methyl-2-Pentanone
Concen: 290.344 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 61324
Ion Ratio Lower Upper
43 100
58 39.6 29.0 43.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#52
Toluene
Concen: 48.419 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion: 92 Resp: 235920
Ion Ratio Lower Upper
92 100
91 173.7 143.8 215.6









#53
t-1,3-Dichloropropene
Concen: 52.206 ug/l
RT: 10.465 min Scan# 1417
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 15104
Ion Ratio Lower Upper
75 100
77 31.8 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#54
cis-1,3-Dichloropropene
Concen: 51.292 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 75 Resp: 161427
Ion Ratio Lower Upper
75 100
77 31.7 24.2 36.4
39 45.3 42.2 63.4





#55
1,1,2-Trichloroethane
Concen: 51.491 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050






Tgt Ion: 97 Resp: 8832


Ion	Ratio	Lower	Upper
97	100		
83	85.6	72.0	108.0
85	54.5	45.4	68.0
99	59.7	50.3	75.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023






#56
Ethyl methacrylate
Concen: 53.680 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion: 69 Resp: 136446

Ion	Ratio	Lower	Upper
69	100		
41	64.9	57.6	86.4
39	34.6	34.1	51.1





#57

1,3-Dichloropropane

Concen: 51.246 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY015391.D

Acq: 31 Aug 2023 09:41

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 151310

Ion Ratio Lower Upper

76 100





78 32.5 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 09/01/2023

Supervised By :Mahesh Dadoda 09/01/2023



#58

2-Chloroethyl Vinyl ether

Concen: 273.857 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. 0.006 min

Lab File: VY015391.D




Acq: 31 Aug 2023 09:41


Tgt Ion: 63 Resp: 248461

Ion Ratio Lower Upper

63 100

106 29.0 21.2 31.8









#59
2-Hexanone
Concen: 315.748 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 50041
Ion Ratio Lower Upper
43 100
58 55.2 25.1 75.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#60
Dibromochloromethane
Concen: 52.222 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion:129 Resp: 104842
Ion Ratio Lower Upper
129 100
127 76.7 38.5 115.5





#61
1,2-Dibromoethane
Concen: 52.640 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:107 Resp: 91490
Ion Ratio Lower Upper
107 100
109 93.9 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#62
4-Bromofluorobenzene
Concen: 50.022 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 95 Resp: 104357
Ion Ratio Lower Upper
95 100
174 84.2 0.0 178.2
176 81.1 0.0 173.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 25076
Ion Ratio Lower Upper
117 100
82 53.4 43.4 65.0
119 31.7 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#64
Tetrachloroethene
Concen: 47.527 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion:164 Resp: 108177
Ion Ratio Lower Upper
164 100
166 126.5 103.0 154.6
129 92.4 77.2 115.8
131 88.5 69.6 104.4





#65
Chlorobenzene
Concen: 47.788 ug/l
RT: 11.514 min Scan# 11590
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 258230
Ion Ratio Lower Upper
112 100
114 31.0 25.5 38.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#66
1,1,1,2-Tetrachloroethane
Concen: 48.481 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion:131 Resp: 95888
Ion Ratio Lower Upper
131 100
133 96.5 47.5 142.6
119 66.0 31.6 94.7









#67
Ethyl Benzene
Concen: 47.158 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 45757
Ion Ratio Lower Upper
91 100
106 31.7 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#68
m/p-Xylenes
Concen: 94.715 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Tgt Ion:106 Resp: 356228
Ion Ratio Lower Upper
106 100
91 199.7 166.1 249.1









#69
o-Xylene
Concen: 47.671 ug/l
RT: 12.032 min Scan# 1674
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 171348
Ion Ratio Lower Upper
106 100
91 212.0 108.1 324.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#70
Styrene
Concen: 49.054 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion:104 Resp: 296466
Ion Ratio Lower Upper
104 100
78 48.7 41.5 62.3
103 55.4 44.6 67.0





#71
Bromoform
Concen: 53.960 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 7444
Ion Ratio Lower Upper
173 100
175 47.5 24.6 73.8
254 0.1 0.2 0.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion:152 Resp: 129122
Ion Ratio Lower Upper
152 100
115 58.5 28.8 86.5
150 173.8 0.0 348.8









#73
Isopropylbenzene
Concen: 46.111 ug/l
RT: 12.331 min Scan# 1723
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:105 Resp: 447222
Ion Ratio Lower Upper
105 100
120 26.8 13.3 39.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#74
N-amy1 acetate
Concen: 54.953 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 43 Resp: 179062
Ion Ratio Lower Upper
43 100
70 43.5 29.7 44.5
55 23.6 18.3 27.5
61 24.7 17.5 26.3





#75
1,1,2,2-Tetrachloroethane
Concen: 53.432 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 123710
Ion Ratio Lower Upper
83 100
131 10.4 5.4 16.2
85 63.7 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#76
1,2,3-Trichloropropane
Concen: 56.470 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 75 Resp: 96694
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 48.301 ug/l
RT: 12.611 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:156 Resp: 110952
Ion Ratio Lower Upper
156 100
77 197.2 92.8 278.6
158 97.0 48.4 145.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#78
n-propylbenzene
Concen: 47.066 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion: 91 Resp: 552356
Ion Ratio Lower Upper
91 100
120 22.7 11.5 34.4





#79
2-Chlorotoluene
Concen: 47.689 ug/l
RT: 12.758 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 313914
Ion Ratio Lower Upper
91 100
126 34.0 17.3 51.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#80
1,3,5-Trimethylbenzene
Concen: 46.851 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion:105 Resp: 380073
Ion Ratio Lower Upper
105 100
120 49.1 23.9 71.7





#81
trans-1,4-Dichloro-2-butene
Concen: 56.149 ug/l
RT: 12.380 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 4358
Ion Ratio Lower Upper
75 100
53 93.2 77.8 116.8
89 46.4 37.4 56.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#82
4-Chlorotoluene
Concen: 47.966 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion: 91 Resp: 328141
Ion Ratio Lower Upper
91 100
126 34.3 17.2 51.6






#83
tert-Butylbenzene
Concen: 45.588 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050






Tgt Ion:119 Resp: 323925
Ion Ratio Lower Upper
119 100
91 64.8 32.9 98.6
134 26.3 13.0 38.9

Manual Integrations


APPROVED

Reviewed By :John Carlone 09/01/2023



Supervised By :Mahesh Dadoda 09/01/2023




#84
1,2,4-Trimethylbenzene
Concen: 47.787 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion:105 Resp: 384905
Ion Ratio Lower Upper
105 100
120 45.2 22.2 66.6





#85
 sec-Butylbenzene
 Concen: 46.171 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY015391.D
 Acq: 31 Aug 2023 09:41

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:105 Resp: 489320
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.2 30.6

Manual Integrations




APPROVED


Reviewed By :John Carlone 09/01/2023
 Supervised By :Mahesh Dadoda 09/01/2023



#86
 p-Isopropyltoluene
 Concen: 46.498 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY015391.D
 Acq: 31 Aug 2023 09:41


Tgt Ion:119 Resp: 413016
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.3 39.8
 91 23.6 12.4 37.4





#87
1,3-Dichlorobenzene
Concen: 47.901 ug/l
RT: 13.367 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 217229
Ion Ratio Lower Upper
146 100
111 40.0 19.4 58.2
148 64.2 31.9 95.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#88
1,4-Dichlorobenzene
Concen: 47.887 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion:146 Resp: 219520
Ion Ratio Lower Upper
146 100
111 38.3 18.9 56.7
148 64.4 32.1 96.5









#89
n-Butylbenzene
Concen: 47.090 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 39368
Ion Ratio Lower Upper
91 100
92 53.2 25.4 76.0
134 26.1 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#90
Hexachloroethane
Concen: 48.533 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Tgt Ion:117 Resp: 77100
Ion Ratio Lower Upper
117 100
201 72.6 41.6 124.8









#91
1,2-Dichlorobenzene
Concen: 49.025 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 20289
Ion Ratio Lower Upper
146 100
111 41.1 19.7 59.0
148 63.7 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 55.724 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41


Tgt Ion: 75 Resp: 25522
Ion Ratio Lower Upper
75 100
155 90.8 41.5 124.6
157 115.3 51.8 155.5





#93
1,2,4-Trichlorobenzene
Concen: 50.142 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:180 Resp: 13032
Ion Ratio Lower Upper
180 100
182 94.5 47.0 141.2
145 30.4 14.5 43.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023




#94
Hexachlorobutadiene
Concen: 46.187 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Tgt Ion:225 Resp: 65458
Ion Ratio Lower Upper
225 100
223 63.0 31.9 95.9
227 63.8 33.0 98.9





#95
Naphthalene
Concen: 55.271 ug/l
RT: 15.239 min Scan# 2200
Delta R.T. 0.006 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:128 Resp: 35584
Ion Ratio Lower Upper
128 100
127 12.9 10.8 16.2
129 10.8 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023



#96
1,2,3-Trichlorobenzene
Concen: 50.787 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. 0.000 min
Lab File: VY015391.D
Acq: 31 Aug 2023 09:41



Tgt Ion:180 Resp: 120400
Ion Ratio Lower Upper
180 100
182 94.8 47.3 141.8
145 31.6 15.4 46.2

