

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090121\
 Data File : VY005927.D
 Acq On : 01 Sep 2021 13:54
 Operator : SY/MD
 Sample : M3597-03
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-PSC124042

Quant Time: Sep 02 07:48:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	1072	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	1762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	1710	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	971	89.884	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 179.760%#	
35) Dibromofluoromethane	7.734	113	383	38.705	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 77.420%	
50) Toluene-d8	10.179	98	2215	53.336	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.680%	
62) 4-Bromofluorobenzene	12.483	95	378	27.107	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 54.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090121\
Data File : VY005927.D
Acq On : 01 Sep 2021 13:54
Operator : SY/MD
Sample : M3597-03
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-PSC124042

Quant Time: Sep 02 07:48:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

