

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090121\
 Data File : VY005932.D
 Acq On : 01 Sep 2021 15:50
 Operator : SY/MD
 Sample : M3612-01
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 M-1

Quant Time: Sep 02 07:49:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	23351	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	33326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	27970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	11915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	10366	44.052	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.100%
35) Dibromofluoromethane	7.722	113	9586	51.219	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.440%
50) Toluene-d8	10.185	98	40289	51.293	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.580%
62) 4-Bromofluorobenzene	12.483	95	10922	41.411	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.820%
Target Compounds						
16) Acetone	3.954	43	1806	30.099	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090121\
Data File : VY005932.D
Acq On : 01 Sep 2021 15:50
Operator : SY/MD
Sample : M3612-01
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
M-1

Quant Time: Sep 02 07:49:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

