

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090123\
 Data File : VY015412.D
 Acq On : 01 Sep 2023 13:57
 Operator : SY/MD
 Sample : 04256-05
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MW-3

Quant Time: Sep 01 23:41:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.789	168	135477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	228879	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	205751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	99843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	87300	60.112	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.220%
35) Dibromofluoromethane	7.716	113	71286	54.074	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.140%
50) Toluene-d8	10.179	98	267466	56.025	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.060%
62) 4-Bromofluorobenzene	12.477	95	91147	51.429	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.860%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	31220	10.957	ug/l	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090123\
Data File : VY015412.D
Acq On : 01 Sep 2023 13:57
Operator : SY/MD
Sample : 04256-05
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MW-3

Quant Time: Sep 01 23:41:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY015412.D
Acq: 01 Sep 2023 13:57




Instrument :
MSVOA_Y
ClientSampleId :
MW-3


Tgt Ion:168 Resp: 135477
Ion Ratio Lower Upper
168 100
99 51.0 43.4 65.0



#20
Methylene Chloride
Concen: 10.957 ug/l
RT: 4.710 min Scan# 474
Delta R.T. 0.006 min
Lab File: VY015412.D
Acq: 01 Sep 2023 13:57

Tgt Ion: 84 Resp: 31220
Ion Ratio Lower Upper
84 100
49 121.4 106.9 160.3
51 35.1 33.2 49.8
86 62.9 51.0 76.6





#33

1,2-Dichloroethane-d4

Concen: 60.112 ug/l

RT: 8.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY015412.D

Acq: 01 Sep 2023 13:57

Instrument :

MSVOA_Y

ClientSampleId :





MW-3

Tgt Ion: 65 Resp: 87300

Ion Ratio Lower Upper

65 100

67 52.4 0.0 101.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY015412.D

Acq: 01 Sep 2023 13:57




Tgt Ion: 114 Resp: 228879


Ion Ratio Lower Upper

114 100

63 19.1 0.0 42.4

88 15.1 0.0 30.6









#35
Dibromofluoromethane
Concen: 54.074 ug/l
RT: 7.716 min Scan# 90
Delta R.T. 0.000 min
Lab File: VY015412.D
Acq: 01 Sep 2023 13:57




Instrument :
MSVOA_Y
ClientSampleId :
MW-3


Tgt Ion:113 Resp: 71286
Ion Ratio Lower Upper
113 100
111 104.3 81.3 121.9
192 20.4 16.9 25.3



#50
Toluene-d8
Concen: 56.025 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. 0.000 min
Lab File: VY015412.D
Acq: 01 Sep 2023 13:57

Tgt Ion: 98 Resp: 267466
Ion Ratio Lower Upper
98 100
100 64.9 50.1 75.1









#62
4-Bromofluorobenzene
Concen: 51.429 ug/l
RT: 12.477 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY015412.D
Acq: 01 Sep 2023 13:57




Instrument :
MSVOA_Y
ClientSampleId :
MW-3


Tgt Ion: 95 Resp: 91147
Ion Ratio Lower Upper
95 100
174 85.8 0.0 178.2
176 83.9 0.0 173.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VY015412.D
Acq: 01 Sep 2023 13:57


Tgt Ion: 117 Resp: 205751
Ion Ratio Lower Upper
117 100
82 52.8 43.4 65.0
119 32.3 26.3 39.5





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.422 min Scan# 1903
 Delta R.T. 0.000 min
 Lab File: VY015412.D
 Acq: 01 Sep 2023 13:57

Instrument :
 MSVOA_Y
 ClientSampleId :
 MW-3



Tgt Ion:	152	Resp:	99843
Ion	Ratio	Lower	Upper
152	100		
115	58.4	28.8	86.5
150	155.2	0.0	348.8

