



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090123\
Data File : VY015404.D
Acq On : 01 Sep 2023 09:26
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Title : SW846 8260
Last Update : Thu Aug 31 20:01:47 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.1	18903	PASS
75	95	30	60	54.0	50885	PASS
95	95	100	100	100.0	94224	PASS
96	95	5	9	6.5	6100	PASS
173	174	0.00	2	1.0	800	PASS
174	95	50	100	86.0	81056	PASS
175	174	5	9	7.5	6068	PASS
176	174	95	101	95.8	77648	PASS
177	176	5	9	6.8	5310	PASS