

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010283.D
 Acq On : 02 Sep 2022 16:54
 Operator : KP/MD
 Sample : N4433-05
 Misc : 4.27g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EP-4

Quant Time: Sep 02 17:29:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	305403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	527083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	491943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	222890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	177810	56.780	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.560%
35) Dibromofluoromethane	7.709	113	158136	45.067	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.140%
50) Toluene-d8	10.178	98	631346	53.742	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.480%
62) 4-Bromofluorobenzene	12.477	95	224692	47.434	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.860%
Target Compounds						Qvalue
20) Methylene Chloride	4.704	84	27766	7.350	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010283.D
Acq On : 02 Sep 2022 16:54
Operator : KP/MD
Sample : N4433-05
Misc : 4.27g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EP-4

Quant Time: Sep 02 17:29:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

