

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010285.D
 Acq On : 02 Sep 2022 17:41
 Operator : KP/MD
 Sample : N4433-02
 Misc : 4.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EP-2

Quant Time: Sep 02 18:04:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	291303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	502199	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	468779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	212153	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	171248	57.331	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.660%
35) Dibromofluoromethane	7.709	113	151111	45.199	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.400%
50) Toluene-d8	10.178	98	599247	53.537	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.080%
62) 4-Bromofluorobenzene	12.477	95	212963	47.148	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010285.D
Acq On : 02 Sep 2022 17:41
Operator : KP/MD
Sample : N4433-02
Misc : 4.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EP-2

Quant Time: Sep 02 18:04:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

