

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010286.D
 Acq On : 02 Sep 2022 18:04
 Operator : KP/MD
 Sample : N4423-01
 Misc : 5.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Sep 02 22:54:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 09/06/2022
 Supervised By :Mahesh Dadoda 09/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	161843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	274905	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	254604	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	113647	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	91591	55.191	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 110.380%			
35) Dibromofluoromethane	7.710	113	83319	45.527	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 91.060%			
50) Toluene-d8	10.179	98	327082	53.382	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 106.760%			
62) 4-Bromofluorobenzene	12.477	95	114993	46.409	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 92.820%			
Target Compounds						Qvalue
11) Tert butyl alcohol	4.942	59	3259m	16.415	ug/l	
20) Methylene Chloride	4.710	84	14558	7.272	ug/l #	83
25) 2-Butanone	6.972	43	5183	7.613	ug/l #	81
30) Chloroform	7.508	83	3761	1.150	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010286.D
Acq On : 02 Sep 2022 18:04
Operator : KP/MD
Sample : N4423-01
Misc : 5.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Sep 02 22:54:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/06/2022
Supervised By :Mahesh Dadoda 09/06/2022

