

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010289.D
 Acq On : 02 Sep 2022 19:14
 Operator : KP/MD
 Sample : N4464-03
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2

Quant Time: Sep 02 22:54:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	290659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	495216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	464420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	218267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	166943	56.014	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.020%
35) Dibromofluoromethane	7.716	113	150500	45.651	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.300%
50) Toluene-d8	10.179	98	593936	53.811	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.620%
62) 4-Bromofluorobenzene	12.477	95	212343	47.754	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.500%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	4.942	59	5077	14.239	ug/l	# 95
20) Methylene Chloride	4.710	84	31696	8.815	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010289.D
Acq On : 02 Sep 2022 19:14
Operator : KP/MD
Sample : N4464-03
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2

Quant Time: Sep 02 22:54:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

