

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010300.D
 Acq On : 03 Sep 2022 01:02
 Operator : KP/MD
 Sample : N4379-05
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-03-9-11

Quant Time: Sep 03 06:10:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	283840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	478603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	448242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	207692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	168354	57.844	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 115.680%			
35) Dibromofluoromethane	7.716	113	146395	45.947	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 91.900%			
50) Toluene-d8	10.179	98	568241	53.270	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 106.540%			
62) 4-Bromofluorobenzene	12.477	95	203677	47.341	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 94.680%			
Target Compounds						
11) Tert butyl alcohol	4.954	59	4095	11.761	ug/l	98
20) Methylene Chloride	4.710	84	44672	12.723	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010300.D
Acq On : 03 Sep 2022 01:02
Operator : KP/MD
Sample : N4379-05
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-03-9-11

Quant Time: Sep 03 06:10:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

