

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010306.D
 Acq On : 03 Sep 2022 03:22
 Operator : KP/MD
 Sample : N4379-12
 Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-06-9.5-11

Quant Time: Sep 03 06:12:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	277926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	467694	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	435406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	200263	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	161990	56.842	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.680%
35) Dibromofluoromethane	7.716	113	142109	45.642	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.280%
50) Toluene-d8	10.179	98	556121	53.349	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.700%
62) 4-Bromofluorobenzene	12.477	95	193517	45.828	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.660%
Target Compounds						
20) Methylene Chloride	4.710	84	31020	9.023	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010306.D
Acq On : 03 Sep 2022 03:22
Operator : KP/MD
Sample : N4379-12
Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-06-9.5-11

Quant Time: Sep 03 06:12:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

