

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010281.D
 Acq On : 02 Sep 2022 16:07
 Operator : KP/MD
 Sample : N4433-03
 Misc : 4.27g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EP-3

Quant Time: Sep 02 16:50:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.783	168	308699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	531526	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	493145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	223997	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	182241	57.574	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.140%
35) Dibromofluoromethane	7.709	113	159169	44.982	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.960%
50) Toluene-d8	10.178	98	637809	53.838	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.680%
62) 4-Bromofluorobenzene	12.483	95	226498	47.413	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.820%
Target Compounds						
20) Methylene Chloride	4.704	84	26100	6.835	ug/l	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010281.D
Acq On : 02 Sep 2022 16:07
Operator : KP/MD
Sample : N4433-03
Misc : 4.27g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EP-3

Quant Time: Sep 02 16:50:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY010281.D
Acq: 02 Sep 2022 16:07




Instrument :
MSVOA_Y
ClientSampleId :
EP-3


Tgt Ion:168 Resp: 308699
Ion Ratio Lower Upper
168 100
99 58.0 46.9 70.3



#20
Methylene Chloride
Concen: 6.835 ug/l
RT: 4.704 min Scan# 473
Delta R.T. -0.006 min
Lab File: VY010281.D
Acq: 02 Sep 2022 16:07

Tgt Ion: 84 Resp: 26100
Ion Ratio Lower Upper
84 100
49 127.0 114.4 171.6
51 38.4 35.3 52.9
86 63.5 53.7 80.5





#33

1,2-Dichloroethane-d4

Concen: 57.574 ug/l

RT: 8.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY010281.D

Acq: 02 Sep 2022 16:07

Instrument :

MSVOA_Y

ClientSampleId :





EP-3

Tgt Ion: 65 Resp: 182241

Ion Ratio Lower Upper

65 100

67 53.6 0.0 99.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VY010281.D

Acq: 02 Sep 2022 16:07




Tgt Ion: 114 Resp: 531526


Ion Ratio Lower Upper

114 100

63 20.1 0.0 43.2

88 15.2 0.0 31.8









#35
Dibromofluoromethane
Concen: 44.982 ug/l
RT: 7.709 min Scan# 91
Delta R.T. -0.001 min
Lab File: VY010281.D
Acq: 02 Sep 2022 16:07




Instrument :
MSVOA_Y
ClientSampleId :
EP-3


Tgt Ion:113 Resp: 159169
Ion Ratio Lower Upper
113 100
111 103.6 82.6 124.0
192 20.6 17.6 26.4



#50
Toluene-d8
Concen: 53.838 ug/l
RT: 10.178 min Scan# 1371
Delta R.T. -0.001 min
Lab File: VY010281.D
Acq: 02 Sep 2022 16:07

Tgt Ion: 98 Resp: 637809
Ion Ratio Lower Upper
98 100
100 65.7 50.2 75.2









#62
4-Bromofluorobenzene
Concen: 47.413 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY010281.D
Acq: 02 Sep 2022 16:07




Instrument :
MSVOA_Y
ClientSampleId :
EP-3


Tgt Ion: 95 Resp: 226498
Ion Ratio Lower Upper
95 100
174 81.2 0.0 177.0
176 79.2 0.0 175.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. -0.001 min
Lab File: VY010281.D
Acq: 02 Sep 2022 16:07

Tgt Ion: 117 Resp: 493145
Ion Ratio Lower Upper
117 100
82 55.0 43.0 64.4
119 31.5 25.4 38.0





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.422 min Scan# 1904
 Delta R.T. -0.006 min
 Lab File: VY010281.D
 Acq: 02 Sep 2022 16:07

Instrument :
 MSVOA_Y
 ClientSampleId :
 EP-3

Tgt Ion:152 Resp: 223997
 Ion Ratio Lower Upper
 152 100
 115 58.4 28.2 84.7
 150 157.2 0.0 347.6

