

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010295.D
 Acq On : 02 Sep 2022 22:42
 Operator : KP/MD
 Sample : VY0902SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0902SBL02

Quant Time: Sep 02 23:12:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	277448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	466523	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	436108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	203020	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	154440	54.286	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.580%
35) Dibromofluoromethane	7.716	113	141883	45.684	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.360%
50) Toluene-d8	10.179	98	556127	53.484	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.960%
62) 4-Bromofluorobenzene	12.477	95	196422	46.760	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010295.D
Acq On : 02 Sep 2022 22:42
Operator : KP/MD
Sample : VY0902SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0902SBL02

Quant Time: Sep 02 23:12:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY010295.D
Acq: 02 Sep 2022 22:42




Instrument :
MSVOA_Y
ClientSampleId :
VY0902SBL02


Tgt Ion:168 Resp: 277448
Ion Ratio Lower Upper
168 100
99 51.5 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 54.286 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY010295.D
Acq: 02 Sep 2022 22:42

Tgt Ion: 65 Resp: 154440
Ion Ratio Lower Upper
65 100
67 53.3 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY010295.D

Acq: 02 Sep 2022 22:42

Instrument :

MSVOA_Y

ClientSampleId :

VY0902SBL02





Tgt Ion:114 Resp: 466523

Ion Ratio Lower Upper

114 100

63 19.6 0.0 43.2

88 15.5 0.0 31.8



#35

Dibromofluoromethane

Concen: 45.684 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.006 min

Lab File: VY010295.D

Acq: 02 Sep 2022 22:42




Tgt Ion:113 Resp: 141883


Ion Ratio Lower Upper

113 100

111 102.7 82.6 124.0

192 20.4 17.6 26.4









#50
Toluene-d8
Concen: 53.484 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY010295.D
Acq: 02 Sep 2022 22:42




Instrument :
MSVOA_Y
ClientSampleId :
VY0902SBL02


Tgt Ion: 98 Resp: 556127
Ion Ratio Lower Upper
98 100
100 66.0 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 46.760 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY010295.D
Acq: 02 Sep 2022 22:42

Tgt Ion: 95 Resp: 196422
Ion Ratio Lower Upper
95 100
174 84.4 0.0 177.0
176 81.3 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. -0.001 min
Lab File: VY010295.D
Acq: 02 Sep 2022 22:42

Instrument :
MSVOA_Y
ClientSampleId :
VY0902SBL02

Tgt Ion:117 Resp: 436108
Ion Ratio Lower Upper
117 100
82 53.4 43.0 64.4
119 32.1 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. -0.006 min
Lab File: VY010295.D
Acq: 02 Sep 2022 22:42

Tgt Ion:152 Resp: 203020
Ion Ratio Lower Upper
152 100
115 57.7 28.2 84.7
150 157.0 0.0 347.6

