

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
 Data File : VY010298.D
 Acq On : 03 Sep 2022 00:15
 Operator : KP/MD
 Sample : N4379-02
 Misc : 6.14g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-12-14

Quant Time: Sep 03 06:09:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.783	168	276472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	471938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	441679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	198566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	164506	58.029	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.060%
35) Dibromofluoromethane	7.710	113	143506	45.677	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.360%
50) Toluene-d8	10.179	98	563051	53.529	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.060%
62) 4-Bromofluorobenzene	12.477	95	199321	46.928	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.860%
Target Compounds						
20) Methylene Chloride	4.710	84	40037	11.707	ug/l	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090222\
Data File : VY010298.D
Acq On : 03 Sep 2022 00:15
Operator : KP/MD
Sample : N4379-02
Misc : 6.14g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-12-14

Quant Time: Sep 03 06:09:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY010298.D
Acq: 03 Sep 2022 00:15




Instrument :
MSVOA_Y
ClientSampleId :
SB-01-12-14


Tgt Ion:168 Resp: 276472
Ion Ratio Lower Upper
168 100
99 54.4 46.9 70.3



#20
Methylene Chloride
Concen: 11.707 ug/l
RT: 4.710 min Scan# 474
Delta R.T. 0.000 min
Lab File: VY010298.D
Acq: 03 Sep 2022 00:15

Tgt Ion: 84 Resp: 40037
Ion Ratio Lower Upper
84 100
49 125.7 114.4 171.6
51 40.3 35.3 52.9
86 64.5 53.7 80.5





#33

1,2-Dichloroethane-d4

Concen: 58.029 ug/l

RT: 8.137 min Scan# 1036

Delta R.T. 0.001 min

Lab File: VY010298.D

Acq: 03 Sep 2022 00:15

Instrument :

MSVOA_Y

ClientSampleId :





SB-01-12-14

Tgt Ion: 65 Resp: 164506

Ion Ratio Lower Upper

65 100

67 52.1 0.0 99.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY010298.D

Acq: 03 Sep 2022 00:15




Tgt Ion: 114 Resp: 471938


Ion Ratio Lower Upper

114 100

63 20.1 0.0 43.2

88 15.2 0.0 31.8









#35
Dibromofluoromethane
Concen: 45.677 ug/l
RT: 7.710 min Scan# 966
Delta R.T. -0.000 min
Lab File: VY010298.D
Acq: 03 Sep 2022 00:15




Instrument :
MSVOA_Y
ClientSampleId :
SB-01-12-14


Tgt Ion:113 Resp: 143506
Ion Ratio Lower Upper
113 100
111 103.2 82.6 124.0
192 20.4 17.6 26.4



#50
Toluene-d8
Concen: 53.529 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY010298.D
Acq: 03 Sep 2022 00:15

Tgt Ion: 98 Resp: 563051
Ion Ratio Lower Upper
98 100
100 64.8 50.2 75.2









#62
4-Bromofluorobenzene
Concen: 46.928 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY010298.D
Acq: 03 Sep 2022 00:15




Instrument :
MSVOA_Y
ClientSampleId :
SB-01-12-14


Tgt Ion: 95 Resp: 199321
Ion Ratio Lower Upper
95 100
174 84.3 0.0 177.0
176 80.9 0.0 175.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. -0.000 min
Lab File: VY010298.D
Acq: 03 Sep 2022 00:15

Tgt Ion: 117 Resp: 441679
Ion Ratio Lower Upper
117 100
82 53.0 43.0 64.4
119 32.3 25.4 38.0





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.422 min Scan# 1903
 Delta R.T. -0.006 min
 Lab File: VY010298.D
 Acq: 03 Sep 2022 00:15

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-12-14

Tgt Ion:152 Resp: 198566
 Ion Ratio Lower Upper
 152 100
 115 58.1 28.2 84.7
 150 158.0 0.0 347.6

