

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090324\
 Data File : VY019420.D
 Acq On : 03 Sep 2024 11:08
 Operator : SY/MD
 Sample : VY0903SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0903SBL01

Quant Time: Sep 04 01:31:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 02:45:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	109343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	194752	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	144149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	56774	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	54530	52.652	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.300%
35) Dibromofluoromethane	7.634	113	60946	49.498	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.000%
50) Toluene-d8	10.109	98	220108	48.561	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.120%
62) 4-Bromofluorobenzene	12.408	95	43447	29.636	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	59.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090324\
Data File : VY019420.D
Acq On : 03 Sep 2024 11:08
Operator : SY/MD
Sample : VY0903SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0903SBL01

Quant Time: Sep 04 01:31:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 02:45:30 2024
Response via : Initial Calibration

