

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090324\
 Data File : VY019434.D
 Acq On : 03 Sep 2024 17:11
 Operator : SY/MD
 Sample : P3812-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-083024

Quant Time: Sep 04 01:36:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 02:45:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	50196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	85288	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	76761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	17931	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	26240	55.190	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.380%
35) Dibromofluoromethane	7.628	113	28567	52.978	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.960%
50) Toluene-d8	10.109	98	93558	47.133	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.260%
62) 4-Bromofluorobenzene	12.408	95	26995	42.047	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	84.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090324\
Data File : VY019434.D
Acq On : 03 Sep 2024 17:11
Operator : SY/MD
Sample : P3812-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-02-083024

Quant Time: Sep 04 01:36:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 02:45:30 2024
Response via : Initial Calibration

