

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090622\
 Data File : VY010358.D
 Acq On : 06 Sep 2022 19:49
 Operator : KP/MD
 Sample : N4500-15
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04-17.5-18.0

Quant Time: Sep 07 05:56:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	9478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	14601	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	12568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	4519	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	4628	47.620	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.240%
35) Dibromofluoromethane	7.716	113	2036	20.946	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	41.900%#
50) Toluene-d8	10.179	98	15614	47.979	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.960%
62) 4-Bromofluorobenzene	12.477	95	4942	36.174	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.340%
Target Compounds						
3) Chloromethane	2.113	50	526	5.092	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090622\
Data File : VY010358.D
Acq On : 06 Sep 2022 19:49
Operator : KP/MD
Sample : N4500-15
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-04-17.5-18.0

Quant Time: Sep 07 05:56:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

